



## Inspection Report

The Ohio State University

OFFICE OF RESPONSIBLE RESEARCH PRACTICES  
300 RESEARCH FOUNDATION BLDG, 1960 KENNY RD  
COLUMBUS, OH 43210

Customer ID: 216

Certificate: 31-R-0014

Site: 001

## OHIO STATE UNIVERSITY CAMPUS

Type: FOCUSED INSPECTION

Date: 08-APR-2025

2.33(b)(4) Critical

## Attending veterinarian and adequate veterinary care.

On February 5, 2025, the death of a pig during an anesthetic event was reported to the Attending Veterinarian and subsequently to the IACUC. The event occurred when the pig was being changed from positive pressure ventilation to spontaneous ventilation. The pop off valve on the anesthetic machine was accidentally left in the closed position. The pig went into cardiac arrest while still under anesthesia and was unable to be resuscitated. Incorrect operation of anesthetic machines can result in injury or death of the patient. Correct by ensuring that personnel providing anesthesia are properly trained on the use of ventilators and anesthetic machines used in the research facility. Following the incident, the veterinary staff reviewed the procedures for ventilation changes with the technical support staff. That review occurred prior to today's inspection.

Correct from April 8, 2025, onward.

This inspection and exit interview were conducted with the Attending Veterinarian.

**Prepared By:** LAUREN MURPHY **USDA, APHIS, Animal Care**  
**Title:** VETERINARY MEDICAL  
OFFICER

**Date:**  
01-JUL-2025

**Received by Title:** Attending Veterinarian

**Date:**  
01-JUL-2025



### Species Inspected

Cust No	Cert No	Site	Site Name	Inspection
216	31-R-0014	001	OHIO STATE UNIVERSITY CAMPUS	08-APR-2025

Count	Scientific Name	Common Name
-------	-----------------	-------------

000000	NONE	NONE
000000	Total	